

Deliver 1 test spot pulse.

Drop initial fluence by 20 percent.

Or reduce by up to 30 percent.

Observe site for tissue changes.

A test spot is mandatory before full treatment.
OBSERVE FOR 4 WEEKS.