



Load the steam sterilizer correctly.

Operate according to validated cycle.

Monitor with physical and chemical indicators.

Support process with maintenance records.

**A failed indicator demands immediate process investigation.
DO NOT RETURN INSTRUMENTS TO SERVICE**