

Wrap entire unit in barrier film.

Cover all exposed electrical charging contacts.

Discard wrap immediately after any puncture.

Units rated IP54 withstand routine wiping.

54
121
134

Lithium components cannot survive autoclave sterilization.

Never exceed standard sterilization heat limits.